

Towards functional 3D T-ray imaging

Bradley Ferguson^{1,2,3}, Shaohong Wang¹, Doug Gray^{2,3}, Derek Abbott²
and X-C Zhang¹

¹ Department of Physics, Applied Physics and Astronomy, Rensselaer Polytechnic Institute, Troy, NY 12180, USA

² Centre for Biomedical Engineering and Department of Electrical and Electronic Engineering, The University of Adelaide, SA 5005, Australia

³ Cooperative Research Centre for Sensor, Signal and Information Processing, Mawson Lakes Boulevard, Mawson Lakes, SA 5095, Australia

E-mail: zhangxc@rpi.edu

Received 5 March 2002

Published 17 October 2002

Online at stacks.iop.org/PMB/47/3735

Abstract

We review the recent development of T-ray computed tomography, a terahertz imaging technique that allows the reconstruction of the three-dimensional refractive index profile of weakly scattering objects. Terahertz pulse imaging is used to obtain images of the target at multiple projection angles and the filtered backprojection algorithm enables the reconstruction of the object's frequency-dependent refractive index. The application of this technique to a biological bone sample and a plastic test structure is demonstrated. The structure of each target is accurately resolved and the frequency-dependent refractive index is determined. The frequency-dependent information may potentially be used to extract functional information from the target, to uniquely identify different materials or to diagnose medical conditions.

1. Introduction

Pulsed terahertz (THz) imaging was first proposed by Hu and Nuss in 1995 (Hu and Nuss 1995). It utilizes the THz, or far-infrared, region of the electromagnetic spectrum and is based upon the technique of terahertz time-domain spectroscopy (THz-TDS). Terahertz imaging has been demonstrated for imaging flames (Mittleman *et al* 1999), scale-model aircraft (Cheville *et al* 1997), leaf moisture content (Hadjiloucas *et al* 1999), skin burn severity, tooth cavities (Ciesla *et al* 2000) and skin tumours (Loffler *et al* 2001). One of the primary advantages of THz imaging is the fact that coherent detection of the broadband pulses provides rich frequency-dependent information on the target's far-infrared optical properties.

Recently, much attention has been focused on three-dimensional (3D) THz imaging techniques. Reflection mode THz tomography systems (Mittleman *et al* 1997) are based on measuring the time-of-flight of reflected pulses. This technique is capable of resolving

the 3D refractive index profile for objects consisting of well-separated layers of differing refractive index. While this technique provides extremely sensitive range resolution of the order of $1\ \mu\text{m}$, the current reconstruction algorithms are only applicable given a number of assumptions that restrict its applicability (Mittleman *et al* 1999). These algorithms neglect multiple reflections, absorption and dispersion in the object to be imaged. These restrictions have motivated research into more general tomographic imaging methodologies.

Other THz imaging algorithms, drawing inspiration from geophysical (Dorney *et al* 2001), radar (McClatchey *et al* 2001) and optical diffraction (Ruffin *et al* 2001) techniques, are capable of mapping the two- and three-dimensional distribution of scattering objects but have currently only been demonstrated for imaging the shape profile of the target object and both the internal structure and the optical properties of the sample are unavailable.

T-ray computed tomography (T-ray CT) offers the potential to extract the frequency-dependent 3D properties of the target (Ferguson *et al* 2002a). This provides a rich four-dimensional dataset that may be used to uniquely identify materials and may potentially have clinical diagnostic applications. In this paper, we demonstrate the extraction of the frequency-dependent refractive index of a 3D structure.

2. Methods

This section describes the hardware and algorithms employed in this work. Section 2.1 describes the chirped probe pulse THz imaging system, which forms the basis of the 3D T-ray CT system. The chirped probe pulse THz imaging technique can provide a potential speed increase of over a factor of 10 over traditional scanned THz imaging systems (Hu and Nuss 1995) and does not require MHz laser repetition rates thus facilitating the use of high power regeneratively amplified lasers. The algorithms employed to reconstruct the frequency-dependent refractive index from T-ray CT data are discussed in section 2.2, and section 2.3 describes the experiments that were performed.

2.1. T-ray imaging with a chirped probe beam

Traditional THz imaging systems (Hu and Nuss 1995) use a standard THz-TDS system with a 20 Hz scanning delay line to accelerate the imaging speed. The THz pulse is focused on the sample and the sample scanned in x and y dimensions to form an image. A number of alternative methods have been suggested. A dramatic improvement in acquisition speed was made using two-dimensional electro-optic detection of a collimated THz pulse (Wu *et al* 1996).

Electro-optic detection of the THz radiation with a chirped probe pulse was first demonstrated by Jiang and Zhang (Jiang and Zhang 1998a). This novel technique allows the full THz waveform to be measured simultaneously rather than requiring a stepped motion stage to scan the temporal profile. This provides a significant reduction in the acquisition time and greatly extends the applicability of THz systems in situations where the sample is dynamic or moving. Indeed, single shot measurements have been demonstrated measuring the transmitted THz response using a single femtosecond light pulse (Jiang and Zhang 1998b).

The hardware schematic of our imaging system is shown in figure 1. The sample was mounted on a x - y translation stage and raster scanned to acquire an image. THz pulses are generated using a regeneratively amplified Ti:sapphire laser (Spectra Physics Hurricane) with an average output power of 700 mW, a pulse duration of 130 fs and a repetition rate of 1 kHz. The pulses are incident on a wide aperture (16 mm) GaAs photoconductive antenna with a bias of 2000 V. The THz radiation is focused by off-axis parabolic mirrors to a spot size of




Figure 1. Simplified hardware schematic used for 2D THz imaging and T-ray CT. The THz pulse is generated using a femtosecond laser pulse incident on a wide aperture antenna on a GaAs substrate with a bias of 2000 V. The optical probe beam is chirped by a grating pair, this allows the full THz temporal profile to be sampled simultaneously via electro-optic sampling. A spectrometer and CCD are used to recover the THz signal.

approximately 1 mm on the target. The transmitted THz pulse is collected with parabolic mirrors and focused onto a 4 mm thick (110) ZnTe electro-optic (EO) detector crystal.

The optical probe pulse is linearly chirped and temporally stretched to a pulse width of 30 ps, using a grating pair with a separation of 4 mm and pitch of 51° . The THz pulse modulates the probe pulse via the EO effect and is recovered with a spectrometer (SPEX 500M) and CCD camera (PI-Pentamax). The CCD exposure time was set to 50 ms. This provided a signal-to-noise ratio (SNR) of approximately 100. The exposure time could be reduced down to 5 ms at the expense of SNR. Normal THz-TDS uses a lock-in amplifier for phase sensitive detection of the THz signal. This gives rise to its impressive SNR performance, which may be as high as 10 000:1 (van Exter and Grischkowsky 1990). A lock-in amplifier is not used in chirped probe sampling and this is one of the disadvantages of this technique, although signal processing algorithms may be used to improve the SNR (Ferguson and Abbott 2001). The bandwidth of the system is limited by the chirped pulse detection technique to approximately 1 THz. The CCD resolution results in a sampling period of 0.15 ps and a spectral resolution of 17 GHz.

2.2. T-ray computed tomography

T-ray computed tomography provides sectional images of objects in an analogous manner to conventional CT techniques such as x-ray CT. The filtered backprojection algorithm is then used to reconstruct the target object including both its 3D structure and its frequency-dependent far-infrared optical properties.

Using the hardware system described in section 2.1, the sample is rotated and a 2D THz image is obtained at multiple projection angles. This technique is quite time consuming; a typical image of size 100×100 pixels measured at 18 projection angles can take over an hour. The imaging speed may be dramatically improved to under a few minutes using 2D THz




Figure 2. The coordinate system used for T-ray CT. The dimensions x , y and z form the standard Cartesian coordinates (z is out of the page). The object is rotated about the z -axis. The THz beam propagates through the target at an angle θ to the x -axis and offset l from the axis of rotation. The inset shows a typical THz pulse as a function of time, t , after propagation through a target.

imaging with a CCD camera (Wu *et al* 1996), however, the diffracted THz radiation then becomes significant and introduces additional processing concerns.

A four-dimensional dataset is acquired in terms of $\{\theta, l, z, t\}$ where θ is the projection angle, l is the horizontal offset from the axis of rotation, z is the vertical dimension parallel to the axis of rotation and t is the time. We desire a reconstruction of the object optical properties in terms of $\{x, y, z, \omega\}$ where x , y and z are the coordinate axes of the object and ω is the frequency. The coordinate system is illustrated in figure 2.

The filtered backprojection algorithm is used to invert the Radon transform to reconstruct the object of interest (Herman 1980). This is based on the approximation that the detected THz signal is simply related to the incident THz pulse by a line integral of the form

$$P_d(\omega, \theta, l) = P_i(\omega) \exp \left[\int_{L(\theta, l)} \frac{-i\omega \hat{n}(r)}{c} dr \right] \quad (1)$$

where P_d is the Fourier transform of the detected THz signal at a frequency ω . P_i is the Fourier component of the incident THz signal at the same frequency, c is the speed of light, L is the straight line between the source and detector and $\hat{n}(r) = n(r) + ik(r)$ is the unknown complex refractive index of the sample.

The filtered backprojection algorithm is given by

$$s(x, y) = \int_0^\pi \left[\int_{-\infty}^\infty S(\theta, \xi) |\xi| \exp[i2\pi \xi l] d\xi \right] d\theta \quad (2)$$

where $S(\theta, \xi)$ is the spatial Fourier transform of the parallel projection data, defined as

$$S(\theta, \xi) = \int_{-\infty}^\infty s(\theta, l) \exp[-i2\pi \xi l] dl \quad (3)$$

where ξ is the spatial frequency in the l direction. $s(\theta, l)$ is extracted from the measured projection data. One of the advantages of T-ray CT over x-ray CT is that s may be one of several parameters derived from the THz pulses. The choice can be made based on the desired application. To perform the reconstruction for the real refractive index $n(\omega)$ at a given frequency, we could use $s = \text{Im}\{-\log[P_d(\omega)/P_i(\omega)]\}$ however this is an oscillatory function and the associated reconstruction is extremely sensitive to noise. A more accurate reconstruction is obtained by using the unwrapped phase of the measured Fourier coefficients such that $s(\theta, l) = \arg\{P_d(\omega)/P_i(\omega)\}$, the filtered backprojection algorithm then reconstructs the phase delay at each pixel on the reconstruction grid, $s(x, y)$. The refractive index is recovered according to

$$n(x, y, \omega) = 1 + \frac{s(x, y)c}{\omega dr} \quad (4)$$

where $dr = dx = dy$ is the reconstruction grid pixel size.

Simpler reconstructions are possible if the sample may be assumed to be non-dispersive. In this case, the bulk refractive index may be reconstructed by choosing s to be the delay of the peak of the THz pulse. Alternatively the bulk absorption coefficient may be reconstructed by using the ratio of the amplitude of the peaks of the detected and incident THz pulses.

2.3. Experimental procedure

Two sample structures were considered: a section of a turkey femur and a plastic structure consisting of a thin plastic cylinder inside a vial. The turkey bone was prepared by soaking it in acetone for 6 h to ensure maximal transmission of the THz radiation. Both samples were imaged using a sampling step of 1 mm in the x and y dimensions to obtain projection images for each 10° angular increment. The bone structure was then reconstructed using the simple non-dispersive algorithms described in section 2.2. The plastic test structure was reconstructed using the non-dispersive methods and using the frequency-dependent algorithm. The timing of the THz pulses was used to reconstruct the structure with high accuracy and then the frequency-dependent reconstruction was performed at all available frequencies. This allowed the frequency-dependent refractive index of the inner plastic cylinder to be extracted and independently verified.

3. Results and discussion

T-ray CT is restricted in biomedical applications by the severe absorption of THz by moist tissue. However some biological tissues, such as bone, have more moderate absorption coefficients, which makes transmission mode imaging feasible. A section of turkey femur was prepared and imaged as described in section 2.3. A frequency-independent reconstruction was performed using the peak amplitude of the THz pulses as the input to the filtered backprojection algorithm. Figure 3(a) shows an optical image of the bone sample and figure 3(b) shows a 3D rendered image produced by combining the reconstructed cross sections at each height, z . The rendered isosurface threshold was constructed by joining the pixels where the reconstructed absorption coefficient fell to 50% of the peak absorption coefficient. This resulted in a reconstructed bone diameter of 18 mm compared to the measured diameter of 22 mm, measured across the widest part of the bone. The irregular surface of the reconstructed image is a result of the surface rendering technique. Variations in the reconstructed absorption as a result of noise or actual variations in the bone's absorption coefficient result in an irregular isosurface. The bone has a fine internal structure that is not accurately reconstructed. This is




Figure 3. A section of turkey femur was imaged with the T-ray CT system. The fine structure inside the bone is of the order of the THz wavelength and therefore causes difficulties in reconstruction. (a) An optical image. The turkey bone was reconstructed and the 3D rendered image generated (b). The reconstruction used the amplitude of the THz pulses at each pixel as the input to the filtered backprojection algorithm.

due to the long wavelength of the THz radiation. Structures with sizes of the order of the wavelength of the THz radiation cause severe diffraction of the THz pulse. Diffraction is not accounted for in the reconstruction algorithm and results in artefacts in the final image. Diffraction tomography techniques may be employed to overcome this limitation (Ferguson *et al* 2002b).

Amplitude reconstructions of the sort shown in figure 3 are useful for revealing structural information, particularly if the internal structure of an object can be revealed. However, a more interesting problem is that of inferring functional information. Recently we have demonstrated classification techniques capable of utilizing the frequency-dependent information provided by THz imaging to differentiate between different tissue types (Ferguson *et al* 2001). The use of such algorithms with T-ray CT data promises to enable functional 3D imaging with T-rays. Biomedical imaging with THz radiation is hindered by the high absorption of most tissues, however, applications may include mapping blood flow and oxidation, or the fat content of tissue. Another promising potential application is in the diagnosis of diseases such as cancer (Loffler *et al* 2001).

Figure 4(a) shows an optical image of the vial containing a thin plastic tube that was used to demonstrate frequency-dependent 3D reconstruction. First the reconstruction was performed using the timing of the peak of the THz pulse in the time domain to yield a reconstruction of the bulk refractive index. The reconstruction of the central slice is shown in figure 4(b) and the 3D rendered image shown in figure 4(c) was produced by combining a number of the reconstructed slices. The isosurface was chosen as the pixels where the reconstructed refractive index was 50% of the maximum index for the outer vial. The reconstructed image dimensions are quite accurate, the vial and cylinder diameters are within 15% of the actual dimensions measured with callipers. However, the vial thickness is much thicker than expected because of the coarse reconstruction grid size of 1.5 mm. The grid size may be improved




Figure 4. A vial and plastic tube were used for testing the T-ray CT system. (a) An optical image of the object. The timing of the peak of the pulse in the time domain was used as the input to the filtered backprojection algorithm and the cross section was reconstructed to yield the refractive index as shown in (b). (c) shows a 3D rendered image produced by combining each of the reconstructed cross sections.




Figure 5. The frequency-dependent refractive index of the plastic inner tube shown in figure 4(a). The refractive index was determined using T-ray CT, which required no assumptions regarding the sample thickness (solid line). The refractive index was also determined with standard THz-TDS using the measured thickness and the algorithm described in (Duvillaret *et al* 1996) (dashed line). The noisiness of the T-ray CT response is due in part to the low SNR of the chirped probe pulse imaging method compared to THz-TDS.

using more projection angles however the resolution is still limited by the wavelength of the THz radiation to approximately 1 mm.

The full frequency-dependent reconstruction algorithm described in section 2.2 was then applied. The central slice was reconstructed at each frequency. The pixels corresponding to the inner tube were averaged to yield the refractive index profile shown in figure 5. For reference the frequency-dependent refractive index of the tube was calculated with normal THz-TDS and is also plotted. There is reasonable agreement between the two techniques although the result from T-ray CT is significantly noisier. This is primarily due to the disparate SNRs of the two measurement techniques.

4. Conclusion

T-ray CT is a powerful extension of terahertz time-domain spectroscopy with several potential applications. We have demonstrated the ability of T-ray CT to extract the frequency-dependent refractive index of a 3D target thereby providing spectroscopic images of weakly scattering objects. T-ray CT provides the refractive index of the sample without requiring *a priori*

knowledge of the sample thickness and allows the internal structure of objects to be revealed. The frequency-dependent reconstruction is noisier than techniques that neglect dispersion and implicitly average the frequency-domain data. However, as the SNR of the T-ray CT hardware is improved it is anticipated that the frequency-dependent information will yield important functional information and enable material classification.

Acknowledgments

This work was supported by the US Army Research Office, the US National Science Foundation and the Australian Research Council. Bradley Ferguson thanks the Australian-American Fulbright Commission.

References

- Chevillat R A, McGowan R W and Grischkowsky D 1997 Late-time target response measured with terahertz impulse ranging *IEEE Trans. Antennas Propag.* **45** 1518–24
- Ciesla C M, Arnone D D, Corchia A, Crawley D, Longbottom C, Linfield E H and Pepper M 2000 Biomedical applications of terahertz pulse imaging *Proc. SPIE* **3934** 73–81
- Dorney T D, Johnson J L, Rudd J V, Baraniuk R G, Symes W W and Mittleman D M 2001 Terahertz reflection imaging using Kirchhoff migration *Opt. Lett.* **26** 1513–5
- Duvillaret L, Garet F and Coutaz J-L 1996 A reliable method for extraction of material parameters in terahertz time-domain spectroscopy *IEEE J. Sel. Top. Quantum Electron.* **2** 739–46
- Ferguson B and Abbott D 2001 Wavelet de-noising of optical terahertz pulse imaging data *J. Fluctuation Noise Lett.* **1** L65–L69
- Ferguson B, Wang S, Gray D, Abbott D and Zhang X-C 2001 Terahertz imaging of biological tissue using a chirped probe pulse *Proc. SPIE* **4591** 172–84
- Ferguson B, Wang S, Gray D, Abbott D and Zhang X-C 2002a T-ray computed tomography *Opt. Lett.* **27** 1312–14
- Ferguson B, Wang S, Gray D, Abbott D and Zhang X-C 2002b T-ray diffraction tomography *OSA Trends in Optics and Photonics (TOPS) The 13th Int. Conf. on Ultrafast Phenomena* vol 72 (Washington, DC: Optical Society of America) pp 450–1
- Hadjiloucas S, Haratzas L S and Bowen J W 1999 Measurements of leaf water content using terahertz radiation *IEEE Trans. Microw. Theory Tech.* **47** 142–9
- Herman G T 1980 *Image Reconstruction From Projections—The Fundamentals of Computerized Tomography* (New York: Academic)
- Hu B B and Nuss M C 1995 Imaging with terahertz waves *Opt. Lett.* **20** 1716–8
- Jiang Z and Zhang X-C 1998a Electro-optic measurement of THz field pulses with a chirped optical beam *Appl. Phys. Lett.* **72** 1945–7
- Jiang Z and Zhang X-C 1998b Single-shot spatiotemporal terahertz field imaging *Opt. Lett.* **23** 1114–6
- Löffler T, Bauer T, Siebert K J, Roskos H G, Fitzgerald A and Czasch S 2001 Terahertz dark-field imaging of biomedical tissue *Opt. Express* **9** 616–21
- McClatchey K, Reiten M T and Chevillat R A 2001 Time resolved synthetic aperture terahertz impulse imaging *Appl. Phys. Lett.* **79** 4485–7
- Mittleman D M, Gupta M, Neelamani R, Baraniuk R G, Rudd J V and Koch M 1999 Recent advances in terahertz imaging *Appl. Phys. B* **68** 1085–94
- Mittleman D M, Hunsche S, Boivin L and Nuss M C 1997 T-ray tomography *Opt. Lett.* **22** 904–6
- Ruffin A B, Decker J, Sanchez-Palencia L, Le Hors L, Whitaker J F, Norris T B and Rudd J V 2001 Time reversal and object reconstruction with single-cycle pulses *Opt. Lett.* **26** 681–3
- van Exter M and Grischkowsky D 1990 Characterization of an optoelectronic terahertz beam system *IEEE Trans. Microw. Theory Tech.* **38** 1684–91
- Wu Q, Hewitt T D and Zhang X-C 1996 Two-dimensional electro-optic imaging of terahertz beams *Appl. Phys. Lett.* **69** 1026–8